



FlexiOH[®]

The Next*Generation*

Orthopedic Immobilization Technology

ULNAR GUTTER IMMOBILIZER

INSTRUCTIONS FOR USE

Please read these instructions carefully before application. Stickers should be used from the label to keep record of product identity.

CONTENTS

- 01. FlexiOH® TECHNOLOGY
- 02. KEY FEATURES
- 03. FlexiOH® ULNAR GUTTER IMMOBILIZER
- 04. INTENDED USE
- 05. DO'S AND DONT'S
- 06. POLICY AND FEEDBACK
- 07. SPECIFICATIONS
- 08. MEASUREMENT GUIDE
- 09. SIZE CHART
- 10. APPLICATION PROCESS
- 11. CONTACT

FlexiOH®

The NextGeneration

Orthopedic Immobilization Technology

FlexiOH® is the first of its kind- an orthopedic immobilizer that has the rigidity to hold the fractured part as well as ensuring proper skin ventilation. FlexiOH® comes as an easy to fix wearable material on the injured part. After wearing, light treatment completes the curing process and makes the FlexiOH® rigid. This process is quick and only takes 3-10 minutes, depending on the size of FlexiOH® immobilizer. After the course of therapy, FlexiOH® can be removed easily without the use of any other external cast cutter or spreader.

The design and the material of FlexiOH® are engineered to provide superior comfort to the patient. A foam, which is in contact with the skin, is bio-compatible (**ISO 10993-10 certified**) and provides cushioning against the hardened polymer. The FlexiOH® immobilizer is designed such a way that the intermittent hollow spaces facilitate circulation of air and evaporation of sweat or water from the skin surface.

Radiolucency of FlexiOH® short arm immobilizer shows that it does not interfere with the radio diagnosis.

KEY FEATURES



Safer



Washable



Latex Free



Comfortable



Breathable



Lightweight



Quick Application

FlexiOH® ULNAR GUTTER IMMOBILIZER

FlexiOH® Ulnar Gutter Immobilizers are designed for immobilization of metacarpals and phalanges. It has an inbuilt soft cushioning padding which is very comfortable for skin contact. It is used to provide support for fingers injuries (ligament instability, sprain or muscle strain), fractures through the neck of the fifth metacarpal. HCP needs to apply it on the patient's hand and then cure it with the RizyCure® light device (recommended) or direct sunlight for around 3-5 minutes depending on the size of immobilizer, until it gets rigid.

FlexiOH® Ulnar Gutter Immobilizer comes in 4 different sizes: S, M L and XL. All of the immobilizers vary in terms of wrist circumference, forearm circumference and finger length. For the best fit of FlexiOH® Ulnar gutter Immobilizer, taking the measurement of the patient's hand is necessary before application.

INTENDED USE

Absolute washable orthopedic Immobilizer for immobilization at meta-carpals and phalanges of the pinky and ring fingers.

- **Boxer's Fracture**
- **Non-displaced fourth or fifth Metacarpal Fractures**
- **Non-displaced fourth or fifth Proximal Phalanx Fracture**
- **Non-displaced fourth or fifth Middle Phalanx Fracture**

CAUTION : In any case of constant discomfort, skin rashes, increased pain, numbness or redness of the skin, Patients should consult immediately to HCPs. FlexiOH® should be removed immediately if required as per HCP's opinion.

DO'S AND DONT'S

Do's

- You can perform normal day-to-day activities which require less movement of the fractured part while wearing FlexiOH® immobilizer.
- You can take bath / shower and enjoy shallow swimming pool but not very deep.
- Clean your immobilizer on regular interval to maintain your skin hygiene.
- Keep injured limb in elevated position in order to reduce swelling. Dry FlexiOH® immobilizer with hair dryer if required.

Dont's

- Do not touch upper surface of FlexiOH® immobilizer during light curing process. Its higher temperature may cause burns.
- Do not lift any weighted object from the fractured hand without your doctors permission.
- Do not try to remove hand from FlexiOH® immobilizer without your doctors permission.
- Avoid swimming/ any outdoor or indoor games which involves strenuous muscle-skeleton activities.
- Protect from hot beverages as it can burn your skin.

POLICY AND FEEDBACK

We aim to provide quality product with patients' comfort and convenience. We are always ready to improve our product. We love to hear from you about your experience with FlexiOH® immobilizer. Your opinion and suggestions are valuable for us.


In case of any inconvenience regarding FlexiOH® immobilizer, you can directly write to us. Contact informations are available on the contact page.

NOTE: Do's and Dont's are just for information purpose. This may not be considered in every fracture cases. Take your physician's advise before following it.

SPECIFICATION

Materials:	Silicone Rubber (USP VI), Light curable resin and Nitrile Rubber foam
Curing System:	RizyCure [®] Light device or Direct sunlight
Ultimate Tensile Strength:	1.25 kN (±10%)
Ultimate Compression Strength:	3.5 kN (±10%)

MEASUREMENT GUIDE



SIZE CHART (in cm)


Size vs Measurement:	S	M	L	XL
Finger Length :	8	8.5	9.5	10
Wrist Circumference:	16	17.5	19.5	21
Forearm Circumference:	21	23.5	28	30

Notes :


- * Above sizes fit better as per dimensions mentioned in the table.
- * However FlexiOH[®] adapts to an oversize to a certain extend by stretchability.
- * We advise healthcare professionals to choose the size as per patient's comfort.

APPLICATION PROCESS


To apply FlexiOH[®] Ulnar Gutter Immobilizer follow the steps shown below :




Take FlexiOH[®] Ulnar Gutter immobilizer from the kit and place it below the hand along with arm as shown in the picture.




Adjust the immobilizer on hand and make sure that it should cover your ring and pinky finger properly. Then start zipping along the arm line.



Now lock the push button to fix it around the palm. Make sure the adjustment should be comfortable to the patient.



After that, adjust the immobilizer as required position.



Take a fully charged RiziCure[®] light device and start curing around the finger area first by keeping it ≤ 1 cm away. Then start curing wrist and forearm area for about 5-10 minutes until it gets rigid. Apply strap loop around fingers to secure the position.

NOTE: Patients may be advised to keep hand position static for about 10 minutes after curing process to achieve its maximum rigidity.

Copy of the design or Technology may infringe IP rights.

in compliance with



North - America :

OrthoHeal LLC

605, Elmwood Park Blvd,
#B2211, River Ridge,
Elmwood, LA 70123. USA

Asia - Pacific :

JC OrthoHeal Private Limited

GF-33/34, Classic Towers, Near Akota Garden,
Akota, Vadodara, Gujarat, India. 390020.

For the complaint and feedback:

Email: contact@orthoheal.com

Phone: +91 265 2350 003

Phone: +1 (201) 565 7633

Web: www.orthoheal.com

